

## TSK-GEL<sup>®</sup> CN-80T<sub>S</sub> Products

<b>Part Numbers:</b>	17348, 4.6mm ID x 15cm, 5 $\mu$ m
	17349, 4.6mm ID x 25cm, 5 $\mu$ m
	17351, 7.8mm ID x 30cm, 10 $\mu$ m

This sheet contains the recommended operating conditions and the specifications for TSK-GEL CN-80T<sub>S</sub> columns and Guardcolumns. Installation instructions and column care information are described in a separate Instruction Manual.

### A. OPERATING CONDITIONS

- Shipping Solvent: 60% Acetonitrile - 40% Water
- Max. Flow Rate: 1.2 mL/min (4.6mm ID)  
  
When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so it doesn't exceed the maximum pressure drop. When changing solvents, use a flow rate equal to 25% of the maximum flow.
- Standard Flow Rate: 0.8 - 1.0 mL/min (4.6mm ID)
- Max. Pressure: 200 kg/cm<sup>2</sup> = 3000 psi (4.6mm ID x 15cm)  
300 kg/cm<sup>2</sup> = 4500 psi (4.6mm ID x 25cm)
- pH Range: 2.0 - 7.5
- Salt Conc. Range: No limitation
- Organic Conc. Range: No limitation
- Temperature: 10 - 50°C. Reduce flow rate when operating below 10°C.
- Cleaning Solvents: (1) High conc. solvent containing organic modifiers\*  
(2) Mixture of organic acids and high conc. organic modifiers\*  
  
\*Acetonitrile and methanol are recommended as a modifier.
- Storage: Store the column in the shipping solvent if it will not be used within three days. Prevent air from entering the column. For overnight storage flush the column with mobile phase at 0.2mL/min.
- Column Protection: An on-line filter (0.2-0.5 $\mu$ m) equipped between pump and injection valve is recommended. Guard columns prevent the column from a contamination of strongly adsorbed solutes. As a general rule, guard columns should be replaced after 30-40 sample injections or when peaks become excessively wide.

### B. SPECIFICATIONS

The performance of TSK-GEL CN-80T<sub>S</sub> column is tested under the conditions described in the data sheet. All columns have passed the following quality control specifications:

- Number of Theoretical Plates (N):  $\geq 11,000$  (15cm columns)  
 $\geq 18,000$  (25 cm columns)
- Asymmetry Factor (AF): 0.7 - 1.8

DS1118 Revised 29AUGUST2006

#### Tosoh Bioscience LLC

156 Keystone Drive  
Montgomeryville, PA 18936  
Phone: (215) 283-5000  
Orders & Technical Service: (800) 366-4875  
FAX: (215) 283-5035



MD Scientific is authorized distributor  
in Denmark for Tosoh Bioscience  
[www.md-scientific.dk](http://www.md-scientific.dk)  
[info@md-scientific.dk](mailto:info@md-scientific.dk)  
Tel. 7027 8565