

## TSK-GEL<sup>®</sup> G2000H<sub>hr</sub> Products

	Column	Corresponding Guard Column
<b>Part Numbers:</b>	17353, G2000H <sub>hr</sub> 7.8mm ID x 30cm, 5 $\mu$ m	17368, Guard column H <sub>hr</sub> -L, 6mm ID x 4cm

This sheet contains the recommended operating conditions and the specifications for TSK-GEL G2000H<sub>HR</sub> columns and guard columns. H-type columns are used exclusively for Gel Permeation Chromatography. Installation instructions and column care information are described in a separate Instruction Manual.

### A. OPERATING CONDITIONS

- Shipping Solvent: Tetrahydrofuran (THF)
- Max./Standard Flow: 2.0 mL/min / 0.5 - 1.0 mL/min
- Max. Pressure: 50 kg/cm<sup>2</sup> = 750 psi
- Multiple Columns: Columns of the same or different pore size are often connected in series to improve resolution and/or to expand the linear portion of the calibration curve. Connect the columns in order of decreasing pore size to avoid overloading from the high MW components. Connect analytical columns using short pieces of 1/16" x 0.01" ID stainless steel tubing.
- Compatible Solvents: benzene, chloroform, toluene, dichloromethane, carbon tetrachloride, o-chlorophenol/chloroform, m-cresol/chloroform, o-dichlorobenzene, dimethylformamide (DMF), dimethylsulfoxide (DMSO), dioxane, n-hexane, hexafluoroisopropanol/chloroform, methylethylketone, N-methylpyrrolidine, acetone, ethanol, 1-chloronaphthalene, trichlorobenzene, decahydronaphthalene (Decalin), methanol, pyridine, quinoline  
  
Important:
  - After the first solvent exchange, exchanges should be limited to similar polarities.
  - Carbon tetrachloride can corrode stainless steel parts in an HPLC system and in the column.
  - How to Change Solvents:
    - Use a linear gradient at a rate of change of 2% per minute.
    - Use a flow rate of  $\leq$  0.5 mL/min for 7.8mm ID columns.
- Temperature: 25° - 140°C
- Sample Size: 0.001 - 0.5mg
- Storage: The column can be left overnight in solvent in the LC system. When it will not be used for longer periods of time, remove the column from the equipment, seal the ends with the provided protective screws, and store it at laboratory temperature. At all times, prevent air from entering the column!
- Column Protection: The use of guard columns is recommended to prolong the life of the analytical column. Guard columns are not for analysis, they do not improve resolution when connected to the main column. They are also not a substitute for filtering the mobile phase and the sample. A guard column does reduce pump pulsation, and further protects the main column by collecting highly adsorptive components and insoluble substances. Guard column life depends greatly on sample cleanliness. As a general rule, guard columns should be replaced when the peaks become excessively wide, or when the peaks show splitting.

### B. SPECIFICATIONS

The performance of TSK-GEL G2000H<sub>HR</sub> columns is tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:

- Number of Theoretical Plates (N): > 16,000
- Asymmetry Factor (AF): 0.7 - 1.6

DS1142 Revised 30AUGUST2006

**Tosoh Bioscience LLC**  
 156 Keystone Drive  
 Montgomeryville, PA 18936  
 Phone: (215) 283-5000  
 Orders & Technical Service: (800) 366-4875  
 FAX: (215) 283-5035



MD Scientific is authorized distributor  
 in Denmark for Tosoh Bioscience  
 www.md-scientific.dk  
 info@md-scientific.dk  
 Tel. 7027 8565