

TSK-GEL® STAT Series Anion Exchange® Products

Part Numbers:	21960, TSKgel Q-STAT, 3mm ID X 3.5cm, 10µm
	21961, TSKgel Q-STAT, 4.6mm ID X 10cm, 7µm
	21962, TSKgel DNA-STAT, 4.6mm ID X 10cm, 5µm

This sheet contains the recommended operating conditions and the specifications for TSK-GEL STAT Series Anion Exchange columns. Installation instructions and column care information are described in a separate Instruction Manual.

A. OPERATING CONDITIONS

- | | |
|------------------------|---|
| 1. Shipping Solvent: | Ion-Exchanged Water |
| 2. Standard Flow Rate: | 1.0 – 2.0mL/min (P/N 21960)
0.5 – 1.4mL/min (P/N 21961)
0.3 – 0.6mL/min (P/N 21962) |
| | When a buffer with high viscosity is used, the maximum flow rate may have to be reduced so as not to exceed the maximum pressure drop. |
| 3. Max. Pressure: | 10Mpa (P/N 21960)
10Mpa (P/N 21961)
15Mpa (P/N 21962) |
| 4. pH Range: | 3.0 - 10.0 (pH above 10 can only be used for a short time) |
| 5. Salt Conc.: | No Limitation |
| 6. Organic Conc.: | <50%. When solvent in column is replaced by distilled or ion-exchanged water, feed the solvent slowly, at flow rates <0.5mL/min (Q-STAT) and <0.25mL/min (DNA-STAT). |
| 7. Temperature: | 10 - 60°C. |
| 8. Cleaning Solvents: | Adsorbed materials can be stripped from the column by repeated injection with one of the following cleaning solutions:

(1) 0.1mol/L NaOH, or
(2) 20 ~ 40% Acetic acid, or
(3) Solution containing aqueous organic solvent such as methanol or acetonitrile, or
(4) Solution containing a solubilizer such as urea and non-ionic surfactants |
| 9. Storage: | Short term storage: keep the column filled with low ionic strength eluent.
For long term storage, replace the solvent in the column with distilled or ion-exchanged water, at flow rates <0.5mL/min (Q-STAT) and <0.25mL/min (DNA-STAT). |

B. SPECIFICATIONS

The performance of TSK-GEL Q-STAT and TSK-GEL DNA-STAT columns are tested under the conditions described in the Data Sheet. All columns have passed the following quality control specifications:

- | | |
|--------------------------------------|--|
| 1. Number of Theoretical Plates (N): | ≥ 200 (PN 21960)
≥ 4,000 (PN 21961)
≥ 4,000 (PN 21962) |
| 2. Asymmetry Factor (AF): | 0.8 - 1.8 (PN 21960)
1.0 - 2.0 (PN 21961)
1.0 - 2.0 (PN 21962) |

DS1232 27October2008

Tosoh Bioscience LLC
156 Keystone Drive
Montgomeryville, PA 18936
Phone: (215) 283-5000
Orders & Technical Service: (800)366-4875
FAX: (215) 283-5035



MD Scientific is authorized distributor in Denmark for Tosoh Bioscience
www.md-scientific.dk
info@md-scientific.dk
Tel. 7027 8565