


Three Micron TSK-GEL ODS-100V Columns for High Throughput LCMS

Introduction

It is well known in HPLC that columns packed with smaller particles are more efficient at the expense of higher back pressure compared to columns packed with larger particles. The gain in efficiency is usually traded off by reducing column length, which in turn results in shorter analysis times and lower column back pressure.

Recently, two lines of 5 micron TSK-GEL ODS-100V and ODS-100Z columns were introduced. Based on chemically inert silica gel, the benefits of these innovative columns have been described in terms of more symmetrical peaks for basic, acidic and chelating compounds. These product lines have now been expanded to include 3 micron packed columns, which are well suited for high throughput LCMS analysis.

Figure 1. Column Efficiency as a Function of Flow Rate for 2mm ID TSK-GEL ODS-100V Columns



Results

Three micron packed columns are now a common tool in many LCMS applications. The data in *Figure 1* clearly demonstrates that from a column efficiency perspective, an acetonitrile-containing mobile phase is preferred over one containing methanol of similar solvent strength. The acetonitrile-containing mobile phase also reduced back pressure by about 40% due to lower eluent viscosity, further enabling faster cycle times in high throughput applications. Standard Reference Material SRM 870 was developed by NIST as a means to classify the many commercially available reversed phase columns into groups of closely-related columns. Amitriptyline, a strong base, and quinizarin, a chelating compound, are included in the mixture together with more traditional compounds. The symmetrical peaks obtained on TSKgel ODS-100V and TSKgel-ODS100Z columns, as shown in *Figure 2*, will simplify mobile phase make up in LCMS and improve the precision of integration.

Figure 2. Column Performance for Basic and Chelating Compounds (SRM 870)





Figure 3 illustrates this further by allowing the use of a low concentration volatile ammonium formate buffer in the analysis of desipramine, an imipramine metabolite.

Figure 3: Desipramine Peak Shape in 5mM Ammonium Formate as a Function of Endcapping



Conclusion

Three micron packed TSK-GEL ODS-100V and ODS-100Z columns are effective tools to increase throughput and improve precision in LCMS applications.



TOSOH

TOSOH BIOSCIENCE

TOSOH Bioscience LLC
156 Keystone Drive
Montgomeryville, PA 18936-9637
Orders & Service: (800) 366-4875
Fax: (215) 283-5035
www.tosohbioscience.com
email: info.sep.am@tosohbioscience.com